



# Certificate of Analysis

Product:

**MabSelect SuRe™ LX**

Code Number:

17547403

Lot No: 10359774

Test/Characteristic:	Limit:	Result:
<b>Breakthrough capacity, <math>Q_{B,10}</math> %</b> mg human IgG / mL resin	min. 55	60
<b>Microbial contamination</b> Colony Forming Units / mL suspension	max. 20	1
<b>Endotoxin activity</b> EU/mL	less than 5.0	<0.25

**Manufactured in compliance with our ISO 9001 certified quality management system.  
No animal derived material is used in raw materials or in the manufacture of this product.**

**Approval Date (Year-Month-Day): 2024-09-25**

**Expiry Date (Year-Month): 2029-09**

**Manufacturing Date (Year-Month): 2024-09**

Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdco LLC or an affiliate

Cytiva Sweden AB  
Björkgatan 30  
SE-751 84 Uppsala, Sweden  
T + 46 (0)18 612 00 00  
Cytiva.com  
Reg.No. SE 55 61 08 1919 01

**Tests and limits according to AS 28985940 Ed. AD**

**Issued (Year-Month-day): 2024-09-25 by Quality Assurance**

**Hemgren, Annika**

This document has been electronically produced and is valid without a signature.